



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/524,800	01/17/2006	Richard Farrar	1781-0017	5890
28078 7590 11/19/2007 MAGINOT, MOORE & BECK, LLP CHASE TOWER 111 MONUMENT CIRCLE SUITE 3250 INDIANAPOLIS, IN 46204			EXAMINER WOODALL, NICHOLAS W	
			ART UNIT 3733	PAPER NUMBER
			MAIL DATE 11/19/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/524,800

Applicant(s)

FARRAR ET AL.

Examiner

Nicholas Woodall

Art Unit

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 September 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-13 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-13 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 07 September 2007 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

1. This action is in response to applicant's amendment received on 09/07/2007.

Drawings

2. The drawings were received on 09/07/2007. These drawings are acceptable.

Allowable Subject Matter

3. The indicated allowability of claim 12 is withdrawn in view of the newly discovered reference(s) to Cusick (U.S. Publication 2005/0021039). Rejections based on the newly cited reference(s) follow.

Claim Rejections - 35 USC § 103

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

5. Claims 1-9 are rejected under 35 U.S.C. 103(a) as being unpatentable over Millard (U.S. Patent 6,712,824) in view of Muller (U.S. Publication 2001/0018589).

Regarding claim 1, Millard discloses a device comprising a guide block including a fixation part and a guide part mounted on the fixation part (see Figure 4 of the reference). The guide part includes at least one structural feature capable of engaging a surgical tool to ensure the tool is located appropriately relative to the tissue of a patient, the guide block includes at least two drives (14 and 16) capable of adjusting the position of the guide part relative to the fixation part in at least two degrees of freedom (column 4 lines 12-24), and the guide part includes at least one position indicator (10) fixed relative

Art Unit: 3733

to the guide part. Millard discloses the device further comprises at least one position monitor capable of tracking the location of the position indicator (column 5 lines 21-25). Regarding claim 2, Millard discloses a device wherein the guide block includes at least three drives capable of adjusting the guide part relative to the fixation part in at least three degrees of freedom (14, 16, and 11; see column 5 lines 38-43). Regarding claim 3, Millard discloses a device wherein the fixation part includes a housing which is hollow and in which the drives are located inside the housing. Regarding claim 4, Millard discloses a device wherein the guide block includes connector shafts which extend from the fixation part to the guide part, which are moved relative to the fixation part by respective drives to cause the location of the guide part to be adjusted. Regarding claim 5, Millard discloses a device wherein the fixation part further includes means for adjusting the drives (5, 8, and 15), which are accessible from outside the housing. Regarding claim 6, Millard discloses a device wherein the structural feature comprises a guide surface capable of being engaged by a blade to define the appropriate alignment for a cut. Regarding claim 7, Millard discloses a device wherein the structural feature comprises an opening capable of receiving a drill bit. Regarding claims 8 and 9, Millard discloses a device wherein the fixation part includes a plurality of openings capable of receiving fasteners to fix the part to the tissue of the patient. Millard fails to disclose the device further comprising a signal generator connected to the drives capable of generating position signals transmitted to the drives to adjust the guide part. Muller teaches a device comprising a signal generator, i.e. a computer, coupled to drives in order to automatically adjust the guides of the device (page 1 paragraph 018). It would

Art Unit: 3733

have been obvious to one having ordinary skill in the art at the time the invention was made to manufacture the device of Mallard further comprising a signal generator coupled to the drives in view of Muller in order to automatically adjust the guide part.

6. Claims 1, 2, 4-6, and 8-13 are rejected under 35 U.S.C. 103(a) as being unpatentable over Pohl (U.S. Patent 4,703,751) in view of Millard (U.S. Patent 6,712,824) and Muller (U.S. Publication 2001/0018589).

Regarding claim 1, Pohl discloses a device comprising a fixation part that is capable of being directly fixed to tissue, a guide part having at least one structural feature capable of engaging a surgical tool, and at least two drives capable of adjusting the guide part relative to the fixation part in at least two degrees of freedom. Regarding claim 2, Pohl discloses a device comprising at least drives for adjusting the position of the guide part relative to the fixation part. Regarding claim 4, Pohl discloses a device further including connector shafts that extend from the fixation part to the guide part and move relative to the fixation part by respective ones of the drives to cause the location of the guide part to be adjusted. Regarding claim 5, Pohl discloses a device wherein the fixation part includes a means for adjusting the drives that are accessible from outside the housing. The applicant invokes 35 U.S.C. 112 6th paragraph in claim 5, which requires the examiner to reference the specification for structure that the applicant designates for performing the function of "adjusting the drives which are accessible from outside the housing". The applicant designates knobs, on page 3 of the specification, as being capable of performing the "means for" functional limitations of claim 5. Pohl discloses knobs being used for a similar function and therefore discloses the limitations

Art Unit: 3733

of claim 5. Regarding claim 6, Pohl discloses a device wherein the guide part comprises a guide surface that is capable of being engaged by a blade. Regarding claims 8 and 9, Pohl discloses a device wherein the guide block has a plurality of openings extending through capable of receiving a fastener. Regarding claim 10, Pohl discloses a device in which at least one of the drives includes a threaded shaft on one of the fixation part and the guide part, and a threaded bore in the other of the fixation part and the guide part in which the threaded shaft can be received by the thread bore and the guide part can be adjusted relative to the fixation part by rotating the threaded shaft relative to the threaded bore. The examiner is interpreting the reference wherein the screw (86) extends through the slot (66) and into a threaded bore (68). By rotating the screw (86), the surgeon is capable of loosening the guide part from the fixation part, so the guide part maybe rotated relative to the fixation part. Once the guide part has been positioned the surgeon can then tighten the screw (86) into the threaded bore (68). Regarding claim 11, Pohl discloses a device wherein at least one of the drives include a knob to cause rotation between the treaded shaft and the threaded bore. Regarding claim 13, Pohl discloses a device wherein at least one of the drives includes a flexible shaft. The examiner is interpreting flexible to mean any degree of flexibility, and all the shafts of the drives are capable of being flexed. Pohl fails to disclose the device further comprising at least one position indicator fixed relative to the guide part with at least one position monitor capable of tracking the location of the position indicator and a signal generator coupled to the drives by electric motors capable of generating position signals transmitted to the drives to adjust the guide part. Millard teaches a device further

Art Unit: 3733

comprising at least one position indicator with at least one position monitor to properly position and orient the device (column 5 lines 21-25). Muller teaches a device further comprising a signal generator, i.e. a computer, coupled to the drives by electric motors, i.e. servo-motors, capable of generating position signals transmitted to the drives to adjust the guide part in order to automatically adjust the guide part (page 1 paragraph 018). It would have been obvious to one having ordinary skill in the art at the time the invention was made to manufacture the device of Pohl further comprising at least one position indicator with at least one position monitor in view of Millard and a signal generator in view of Muller in order to properly position and orient the device and to automatically adjust the guide part.

Response to Arguments

7. Applicant's arguments with respect to claims 1-13 have been considered but are moot in view of the new ground(s) of rejection. The examiner has presented new grounds of rejection which were necessitated by the amendments of the claims, but the previously indicated allowability of claim 12 has been withdrawn making this office action non-final.

Conclusion

8. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. See PTO-892 for cited references the examiner felt were relevant to the application.


Art Unit: 3733

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Nicholas Woodall whose telephone number is 571-272-5204. The examiner can normally be reached on Monday to Friday 8:00 to 5:30 EST..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

NWWW



EDUARDO C. ROBERT
SUPERVISOR, PATENT EXAMINER